



The Future of Healthcare



RESULT in  
20 min  
With EDTA Tube



ESR test performed on same tubes in which blood samples taken for CBC (EDTA for CBC)

Test can be done on small volume of whole blood (1.5ml)




NOT REQUIRED  
CITRATE TUBE

ESR test results on a compact 10"/7" touch screen, designed for seamless interaction and clear visualization



- User & Environment friendly.
- Keep on loading EDTA tube samples randomly & result in 20 minutes.
- Prevent waste accumulation and reduce the cost of liquid waste treatment.

...  
...  
...



RESULT CALCULATION  
BY USING MODIFIED  
WESTERGREN METHOD

## Technical Specifications

Power supply	110-230V AC 50Hz/60Hz
Dimension	189mm(L) x 282mm(W) x 23mm(H)
Weight	19°C-30°C
Operating temperature	20 to 80% without condensation.
Relative humidity	10°C-40°C
Storage conditions	Vio 185-232, HL7 protocol
LIS interface	2 USB port for printer & barcode reader
USB port	10
Reading channels	ISO 13485, USFDA registered, CDSCO registered, CE certified
Standards & Regulations	IEC 61010-1:2010/AMD :2016, HC 61010-1201), 6C 6K10-2-1012018 IEC 61326-1:2020, IEC 61326-2-6:2020

## Product Code

Instrument	Heme ESR	HEME ESR 10 CHANNEL	
Consumables	Heme ESR Recharge	RECHARGE TEST 500 Nos	
Accessories (Optional)	HEME-BCR BARCODE READER	HEME-TP THERMAL PRINTER	HEME-TPR THERMAL PAPER

Marketed By

**HEME DIAMED LLP**  
2ND FLOOR,4/1981, ANJIKATH LANE NEAR VALLATHOL JUNCTION  
VAZHAKKALA, THRIKKAKARA - P.O. -682021 KERALA

## Contact us

Tel: +917395966499  
Email:info@hemedamed.com  
Web: www.hemedamed.com